

Eaton 10000 Woodward Avenue Woodridge, IL 60517 USA

## **UK DECLARATION OF CONFORMITY**

We, Eaton Corporation, declare, under our sole responsibility, as the manufacturer that the below product(s) comply with the following regulation(s), provided that it is installed, maintained and used in the application intended for, with respect to the relevant manufacturer's instructions, installation standards and "good engineering practices".

Type of equipment:	MEDICAL UPS
Brand:	Eaton, Eaton Tripp Lite Series, Eaton
Model number(s):	SMX700HGL SMX1200XLHGL
Series number(s):	AG-031B AG-031C
Conform with the following standards:	EN 60601-1:2006+A1:2013+A2:2021 EN 62040-1:2008+A1:2013 EN/IEC62040-2:2018 EN60601-1-2:2015+A1:2021 IEC60601-1-2:2020 (4.1 Ed) CISPR 11:2015+AMD1:2016+AMD2:2019 EN IEC 63000:2018
Product(s) here with complies with the requirements of the following UK legislation:	Medical Equipment Regulations 2016 (Medical) Electrical Equipment Safety Regulations 2016 (Safety) Electromagnetic Compatibility Regulations 2016 (EMC) The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 (RoHS)
Manufacturer Authorised Representative in the UK:	Eaton PO Box 554, Abbey Park Southampton Rd Titchfield, PO14 4QA, United Kingdom

Cheryle Chavarria Cheryle Chavarria Regulatory Team Leader Eaton, Woodridge

933FFB RevD, Issued by Eaton 10/16/2024 Template 933F40 RevH, November 2023