

## UK CA

## **UK DECLARATION OF CONFORMITY**

We/Manufacturer:

Tripp Lite 1111 W. 35<sup>th</sup> Street, Chicago, Illinois 60609, USA

as manufacturers declare, under our sole responsibility, that the below product(s) comply with the regulations to declare conformity.

Type of equipment:	Medical UPS
Model number(s):	HCINT350S1, HCINT350SNR, HCINT350S1-60HZ
Series number(s):	AGSM5508
Conform with the following standards:	EN 60601-1 EN 55011:2009+A1:2010 EN 60601-1-2:2015 EN IEC 63000:2018
Product(s) herewith complies with the requirements of the following UK legislation:	Electrical Equipment Safety Regulations 2016 (Safety)
	Electromagnetic Compatibility Regulations 2016 (EMC)
	The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 (RoHS)
Manufacturer Representative in the UK:	Eaton PO Box 554, Abbey Park Southampton Rd Titchfield, PO14 4QA, United Kingdom

John DeCicco Engineering Manager, Tripp Lite, Chicago John DeCicco

> 934057 RevA, Issued by Tripp Lite 6/14/2022 Template 933F40 RevE, April 2022