



Eaton
10000 Woodward Avenue
Woodridge, IL 60517
USA

UK DECLARATION OF CONFORMITY

We, Eaton Corporation, declare, under our sole responsibility, as the manufacturer that the below product(s) comply with the following regulation(s), provided that it is installed, maintained and used in the application intended for, with respect to the relevant manufacturer's instructions, installation standards and "good engineering practices".

Type of equipment:	Medical UPS
Brand:	Tripp Lite, Eaton Tripp Lite Series, Eaton
Model number(s):	HCINT350S1, HCINT350S1-60Hz, HCINT350SNR
Series number(s):	AGSM5508
Conform with the following standards:	EN IEC 60601-1:2005+A2:2020 EN 55011:2016+A1:2017+A2:2021+A11:2020 IEC 60601-1-2:2014+A1:2020 EN 60601-1-2:2015+A1:2021 EN IEC 63000:2018
Product(s) here with complies with the requirements of the following UK legislation:	Medical Equipment Regulations 2016 (Medical) Electromagnetic Compatibility Regulations 2016 (EMC) The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 (RoHS)
Manufacturer Authorised Representative in the UK:	Eaton PO Box 554, Abbey Park Southampton Rd Titchfield, PO14 4QA, United Kingdom

Cheryle Chavarria
Regulatory – Team Leader
Eaton, Woodridge