



1111 W. 35th Street, Chicago IL 60609 Tel: 773-869-1111 www.tripplite.com

## **Declaration of Conformity**

The following product(s) below comply with the listed Directives of the European Parliament and of the Council of the restriction of use of certain hazardous substances in electrical and electronic equipment (also known as "RoHS Recast").

Medical UPS Type of Equipment:

HCINT350S1 Model Number(s):

> HCINT350S1-60Hz HCINT350SR

Series Number(s): AGSM5508

Year of CE Marking (Reissue): 2018

Applicable Council Directive(s): LVD 2014/35/EU

EMC 2014/30/EU

RoHS2 2011/65/EU as amended by 2015/863/EU

Standards to which conformity FN 60601-1-2:2015

is being declared: EN 55011:2009+A1:2010

## **RoHS Exemptions:**

6(a) Lead as an alloying element in steel containing up to 0.35% lead by weight.

6(b) Lead as an alloying element in aluminum containing up to 0.4% lead by weight.

6(c) Lead as an alloying element in copper containing up to 4% lead by weight.

7(c)-II Lead in dielectric ceramic in capacitors for a voltage of 125VAC or 250VDC or higher

8(b) Cadmium and its compounds in electrical contacts

Tripp Lite Document No. 953341 RevD Issued by Tripp Regulatory; 11/26/2018



